



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

Eclipse Aesthetics, LLC
% Mr. Tom O'Brien
Chief Executive Officer
13998 Diplomat Drive
Dallas, Texas 75234

Letter dated: January 10, 2013

Re: K122922

Trade/Device Name: Tri-Beam Nd: YAG Laser System

Regulation Number: 21 CFR 878.4810

Regulation Name: Laser surgical instrument for use in general and plastic surgery and in dermatology

Regulatory Class: Class II

Product Code: GEX

Dated: December 05, 2012

Received: December 11, 2012

Dear Ms. O'Brien:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA).

You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical

device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Mark N. Melkerson

Mark N. Melkerson
Acting Director
Division of Surgical Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K122922

Device Name: Tri-Beam Nd:YAG Laser System

Indications for Use:

The Tri-Beam Nd:YAG Laser System is indicated for: the incision, excision, ablation, vaporization of soft tissues for general dermatology, dermatologic and general surgical procedures for coagulation and hemostasis.

532nm Wavelength (nominal delivered energy of 585 nm and 650 nm with optional dye handpieces):

- Tattoo removal: light ink (red, tan, purple, orange, sky blue, green)
- Removal of Epidermal Pigmented Lesions
- Removal of Minor Vascular Lesions including but not limited to telangiectasias
- Treatment of Lentiginosities
- Treatment of Café-Au-Lait
- Treatment of Seborrheic Keratoses
- Treatment of Post Inflammatory Hyper-Pigmentation
- Treatment of Becker's Nevi, Freckles and Nevi Spilus

1064nm Wavelength:

- Tattoo removal: dark ink (black, blue and brown)
- Removal of Nevus of Ota
- Removal or lightening of unwanted hair with or without adjuvant preparation.
- Treatment of Common Nevi
- Skin resurfacing procedures for the treatment of acne scars and wrinkles

Prescription Use xx
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Neil R Ogden

2013.01.04 14:51:12 -05'00'

(Division Sign-Off) for MXM

Division of Surgical Devices

510(k) Number K122922